



Product Correction

Urgent - Immediate Action Required

Date Issued

December 20, 2018

Product

Product Name	List Number	Serial Number	UDI
Alinity i Processing Module	03R65-01	Ai01001 to Ai01455 Ai01457, Ai01459	Not Applicable

Explanation

Abbott has identified an issue with the Alinity i Gear Pump Assembly (part number A-30108552-01) resulting in foaming/bubbling out of the bottle reservoir for Concentrated Wash Buffer and an unexpected amount of dried residue of buffer. The foaming/bubbling is a result of an improper calibration of the pump's operating speed.

Patient Impact

The Concentrated Wash Buffer (List Number 06P13) contains 5-Bromo-5-nitro-1,3-dioxane, which may produce an allergic reaction when in contact with skin.

Necessary Actions

Please continue to follow the instructions provided in the Alinity ci-series Operations Manual when accessing the bulk solution reservoir area to prevent skin contact.

Your Abbott representative will be contacting you to schedule the replacement of the Gear Pump Assembly A-30108552-01 in the Concentrated Wash Buffer position with a properly calibrated pump part number A-30111949-01.

Please complete and return the attached Customer Reply form.

If you have forwarded the product listed above to other laboratories, please inform them of this Product Correction and provide to them a copy of this letter.

Please retain this letter for your laboratory records.

Contact Information

We sincerely regret any inconvenience this may have caused your laboratory.

If you or any of the health care providers you serve have any questions regarding this information, U.S. Customers please contact Customer Service at 1-877-4ABBOTT (available 24 hours a day, 7 days a week). Customers outside the U.S., please contact your local area Customer Service.

**Contact
Information -
continued**

Adverse reactions or quality problems experienced with the use of this product may be reported to the FDA's MedWatch Adverse Event Reporting program either online (<http://www.fda.gov/MedWatch/report.htm>), by mail (<http://www.fda.gov/MedWatch/getforms.htm>), by phone (1-800-332-1088), or by fax (1-800-FDA-0178).

If you have experienced any patient or user injury associated with this Field Action, please immediately report the event to your local area Customer Service.
