



Product Recall

Immediate Action Required

Date Issued April 25, 2016

Explanation The purpose of this letter is to inform you of a product recall and reduced expiration dating for select ARCHITECT STAT High Sensitive Troponin-I Reagent lots listed in Table 1 below and to provide instructions on what actions your laboratory must take.

Current in-date reagent lots per Table 2 will remain at their current expiration date, therefore no further action is required for these current lots.

Table 1:

Product

Product Description	Kit Size	List Number	Lot Number	Expiration Date	Necessary Actions
ARCHITECT STAT High Sensitive Troponin-I Reagent Kit	100 Test Kit	3P25-25	60103UI00	30 Jul 2016	Once an alternate reagent lot is available, destroy any remaining inventory of these impacted lots, Refer to Necessary Actions, Section A Table 3 for further details on actions required.
		3P25-26	55946UI00	09 Nov 2016	
			58927UI00	28 Feb 2017	
			58927UI01	28 Feb 2017	
			63010UI00	18 Nov 2016	
	500 Test Kit	3P25-35	59259UI00	30 Jul 2016	Refer to Necessary Actions, Section A Table 3 for further details on actions required.
		3P25-36	55957UI00	09 Nov 2016	
			59904UI00	28 Feb 2017	
			62027UI00	18 Nov 2016	
	100 Test Kit	3P25-25	62019UI00	26 Nov 2016	Change the expiration date of these impacted lots Refer to Necessary Actions, Section B Table 4 for further details on actions required.
			62021UI00	04 Nov 2016	
			63139UI00	07 Dec 2016	
		3P25-26	62025UI00	04 Oct 2016	
			62026UI00	07 Dec 2016	
			62354UI00	07 Dec 2016	
	500 Test Kit	3P25-35	62022UI00	04 Nov 2016	
			63108UI00	26 Nov 2016	
			63137UI00	07 Dec 2016	
		3P25-36	62255UI00	05 Oct 2016	
			63051UI00	07 Dec 2016	

Table 2: Current in date reagent lots will remain at their current expiration date.

Product Description	Kit Size	List Number	Lot Number	Expiration Date
ARCHITECT STAT High Sensitive Troponin-I Reagent Kit	100 Test Kit	3P25-25	58918UI00	08 June 2016
			59900UI00	16 July 2016
	500 Test Kit	3P25-35	57943UI00	08 June 2016
			59903UI00	16 July 2016

Predictive quality assurance monitoring of ARCHITECT STAT High Sensitive Troponin-I reagent lots was initiated following the field action that occurred in January (FA21JAN2016). As a result, Abbott has identified a potential for increased calibration failures prior to the expiration date.

The root cause for this situation is under investigation for an appropriate long-term corrective action.

Patient Impact

There is no impact to generated patient results. There is a potential for delay in results due to calibration errors and failure to calibrate.

Necessary Actions

We direct you to follow the Necessary Actions below to reduce the possibility of the incidence of calibration errors.

A. Once an alternate reagent lot is available, destroy any remaining inventory of the following ARCHITECT STAT High Sensitive Troponin-I reagent lots.

Table 3:

Kit Size	List Number	Lot Number
100 Test Kit	3P25-25	60103UI00
		55946UI00
	3P25-26	58927UI00
		58927UI01
		63010UI00
500 Test Kit	3P25-35	59259UI00
		55957UI00
	3P25-36	59904UI00
		62027UI00
If...	Then...	
You do NOT have an alternate reagent lot available in inventory and you have generated a valid calibration curve	<p>Immediately order a replacement reagent lot. Continue to use these lots and follow the recommended QC procedures in the Package Insert. These lots have an increased potential for calibration failures with continued use. Should repeat calibration failures occur, it would prevent you from being able to generate patient results.</p> <p>Once you receive the replacement reagent lot, destroy any remaining inventory of the reagent lots according to your laboratory procedures.</p>	
You HAVE an alternate reagent lot available in inventory	<p>Discontinue use of these lots immediately and switch to the alternate reagent lot.</p> <p>Destroy any remaining inventory of the reagent lots according to your laboratory procedures.</p>	

**Necessary
Actions
(cont.)**

- Your local Abbott representative can help provide you with replacement product, financial credit and/or recommended alternative solutions.
- All future lots will carry the appropriate dating prior to delivery.

B. Adjust the expiration date of the following ARCHITECT STAT High Sensitive Troponin-I reagents lots

Table 4:

Kit Size	List Number	Lot number	From Expiration Date...	To Expiration Date...
100 Test Kit	3P25-25	62019UI00	26 Nov 2016	07JUL2016
		62021UI00	04 Nov 2016	04JUN2016
		63139UI00	07 Dec 2016	07JUL2016
	3P25-26	62025UI00	04 Oct 2016	02JUL2016
		62026UI00	07 Dec 2016	07JUL2016
		62354UI00	07 Dec 2016	07JUL2016
500 Test Kit	3P25-35	62022UI00	04 Nov 2016	04JUN2016
		63108UI00	26 Nov 2016	07JUL2016
		63137UI00	07 Dec 2016	07JUL2016
	3P25-36	62255UI00	05 Oct 2016	02JUL2016
		63051UI00	07 Dec 2016	07JUL2016
If...			Then...	
You are currently using or have inventory of these lots			Please adjust the expiration date as instructed above. As this is a manual change in expiration date, the ARCHITECT software will continue to track to original dating as assigned during the manufacture of these reagents. Manual tracking of the newly assigned expiration dates will have to be performed for each individual ARCHITECT instrument, kit and laboratory.	

C. Complete and return the Customer Reply Form.

D. If you have forwarded any of the products listed within this letter to other laboratories, please inform them of this Product Recall and provide them a copy of this letter.

E. Please retain this letter for your laboratory records.

F. Although Abbott has not yet identified a root cause; actions are being taken to mitigate and/or prevent any further reduction of product expiration dating and maintain product supply.

**Necessary
Actions
(cont.)**

These actions include:

- Internal testing of all on-market lots will continue to monitor performance.
- Until root cause is determined and corrective actions are implemented, new reagent lots will have a proactively shortened expiration date of 4 months from date of manufacture. The new expiration date will require adjustments to your laboratory inventory and order management practices.
Please contact your local Abbott representative for assistance and advice on optimization of workflow in your laboratory.

**Contact
Information**

We regret the inconvenience this situation is causing you. If you or any of the health care providers you serve have any questions regarding this information, please contact your local Abbott representative.
