



Product Recall

Urgent - Immediate Action Required

Date Issued

December 18, 2018

Product

Product Description	List Number (LN)	Lot Number	UDI
ARCHITECT B12 Reagent Kit	7K61-35	91345UI00	N/A

Explanation

The purpose of this letter is to inform you of a product recall for ARCHITECT B12 Reagent Kit, LN 7K61-35, lot 91345UI00 and to provide instructions on what actions your laboratory must take.

Abbott has identified that the incorrect components, 2 assay diluent bottles (7K61J) or 2 pre-treatment 1 bottles (7K61U), may be present in a portion of batch 91345UI00.

Kits that do not contain all correct components cannot be scanned on the ARCHITECT instrument, an error code will be generated upon loading on the reagent inventory screen (eg: Error code 0900 or 0201).


The root cause for this situation is under investigation for an appropriate corrective action.

Patient Impact

There is no impact to generated patient results as the instrument will prevent use of a kit with the incorrect components present.


Necessary Actions

1. Inspect each kit of lot 91345UI00 in inventory before use.
 - Confirm the following components are present as detailed in the IFU
 - Microparticles – 7K61G




Necessary
Actions
Continued


▪ Conjugate -7K61H




▪ Assay Diluent – 7K61J



▪ Pre-Treatment reagent 1 – 7K61U



▪ Pre-Treatment Reagent 2 – 7K61Y



▪ Pre-Treatment Reagent 3 – 7K61X



**Necessary
Actions
continued**

- If the correct components are present, continue to use as normal.
 - If all correct components are not present discard.
2. Your local Abbott representative can help provide you with replacement product and/or credit based on the total number of kits destroyed.
 3. Complete and Return the Customer Reply Form
 4. If you have forwarded the product listed above to other laboratories, please inform them of this Product Recall and provide to them a copy of this letter.
 5. Please retain this letter for your laboratory records.

**Contact
Information**

We regret the inconvenience this situation is causing you. If you or any of the health care providers you serve have any questions regarding this information, please contact your local Abbott representative.

If you have experienced any patient or user injury associated with this Field Action, please immediately report the event to your local area Customer Service.
